

Test Code	Test Description	Bare minimum volume (mL)	Acceptable/standard volume (mL)	Ideal volume (mL)
W04	Ammonium-N:	5	20	20
W05	Nitrate/Nitrite-N (0.02 to 5.00 mg/L)	5	20	20
W06	Nitrate/Nitrite-N (0.5 to 20.0 mg/L)	5	20	20
W07	Nitrite-N:	5	20	20
W08	Orthophosphate-P (0.02 to 2.00 mg/L)	5	20	20
W09	Orthophosphate-P (0.20 to 20.0 mg/L)	5	20	20
W10	Total phosphorus (TP):	20	50	100
W11	Chloride:	5	20	20
W12	Total Organic Carbon (TOC)	25	50	100
W13	Alkalinity:	50	100	100
W14	pH:	20	40	100
W15	Solids (total, total dissolved, total suspended)	100	250	250
W16	Solids (total volatile, suspended volatile)	100	250	250
W17	Specific conductance	10	20	50
W18	Ion Chromatography:	1	10	10
W19	ICP (15-element)	10	20	20
W20	ICP (27-element)	10	20	20
W21	Water Digest	25	50	100